


# Multi-Use PDF TEMP&RH Data Logger

## User Manual



## Product Introduction

The device is mainly used to monitor the temperature and humidity of food, medicine, chemical products and other products during storage and transportation. It is widely used in all links of warehousing and cold chain logistics, such as reefer containers, refrigerated trucks, refrigerated distribution boxes, cold storage laboratory, etc. Compact, light and economical to use. After the recording is completed, inserting it into the USB port of PC or computer, it will automatically generate reports without any driver.

## Product Features

- **Designed to be multi-use**
- **Temperature & Humidity measurement and recording**
- **Wide measuring range, high accuracy and large data memory**
- **Statistics available on LCD screen**
- **No software needed to export data**
- **Automatically generates PDF report and CSV file**
- **Programmable software for logging parameters, alarms, and start delay**

## Technical Parameter

Item	Parameter
Temperature Scale	°C or °F
Accuracy	±3%RH; ±0.5°C (-20°C ~ +40°C), ±1.0°C (other)
Measuring Range	Humidity 0% ~ 100%RH, Temp -30°C ~ 60°C
Resolution	0.1%RH typically, 0.1°C
Memory Capacity	12,000
Recording Options	Delayed or Push-To-Start Logging
Interval	10 seconds to 18 hours adjustable [Default: 10 mins]
Start Delay	Programmable (0~254 mins) [Default: 30 mins]
Alarm Delay	Programmable (0~960 mins) [Default: 10 mins]
Alarm Range	Programmable high or low alarm limit for each channel [Default: <2°C or >8°C, <40%RH or >80%RH]
Shelf Life / Battery	Typically 1 Year; CR2032 3.0V Lithium Battery (Depending on sampling rate and environment)
Report Generation	Simultaneously generate PDF report and CSV file
Communication Interface	USB 2.0
Time Zone	UTC +0:00 (Default)
Dimensions	89mm*36mm*16mm
Product Weight	25g

## Operating Instruction

### a. Start Recording

Press and hold the “▶” button for more than 3s until the “OK” light is on, and the “▶” or “WAIT” displays on the screen, indicating starting recording successfully.


### b. Mark

When the device is recording, press and hold the “▶” button for more than 3s, and the screen will switch to the “MARK” interface. The number of “MARK” will increase by one, indicating data was marked successfully.  
(Note: One record interval only can mark one time, the logger can mark 6 times in one recording trip. Under status of start delay, the mark operation is disabled.)

### c. Stop Recording

Press and hold the “■” button for more than 3s until the “ALARM” light is on, and the “■” displays on the screen, indicating stopping recording successfully.  
(Note: If the logger is stopped during the status of start delay, a PDF report generated when inserted into PC but without data.)

## LCD Display Instruction




During normal recording process, shortly press “▶” to switch to different display interface. The interfaces shown in sequence are respectively:


Real-time Temperature Interface → Real-time Humidity Interface → LOG Interface → MARK Interface → Temperature Upper Limit Interface → Temperature Lower Limit Interface → Humidity Upper Limit Interface → Humidity Lower Limit Interface.

### Note:


- If the device is used for the first time or after re-configuration, the real-time temperature interface will be the initialization interface.
- Real-time temperature and humidity interface is updated every 10s.




① Real-time Temperature Interface  
(Initialization)



② Real-time Humidity Interface  
(Initialization)



Real-time Temperature Interface  
(Above Upper Limit)



Real-time Humidity Interface  
(Below Lower Limit)

## At the real time temperature / humidity interface:

If the symbol “►” lights, indicating the data logger is recording;


If the symbol “■” lights, indicating the data logger has stopped recording;

If the “WAIT” is lit, indicating the data logger is in the status of start delay;


If the symbol “√” lights, it means that the temperature and humidity are within the normal range;

If the symbol “x” and “↑” light, indicating the measured temperature/humidity exceeds its temperature/humidity upper limit;


If the symbol “x” and “↓” light, indicating the measured temperature/humidity exceeds its temperature/humidity lower limit.




③LOG Interface  
(Stop recording status)




④MARK Interface  
(Stop recording status)




⑤Temperature: Above Upper Limit  
(Stop recording status)



⑥Temperature: Below Lower Limit  
(Stop recording status)



⑦Humidity: Above Upper Interface  
(Stop recording status)







⑧Humidity: Below Lower Limit  
(Stop recording status)

## LED Indicator Description

Status	Operation	Indicator status
Inactivated	Short press the button	The “OK” ( green light) and the “Alarm” (red light) will flash simultaneously.
Activated	<p>During the recording process, the device will automatically flash every 10s.</p> <p>If the “OK”(green light) flashes every 10s, it means that there is not over temperature/humidity during the recording process.</p> <p>If the “Alarm” (red light) flashes every 10s, it means there is over temperature/humidity during the recording process.</p> <p>Note: As long as over temperature is found during the recording process, the “OK” light will no longer flash.</p>	
	Not press the button	The “OK” (green light) and “Alarm”(red light) don’t flash.
Stopped	Short press the button	The “OK” (green light) flashes, indicating there is not over temperature during the recording process.
		The “Alarm” (red light) flashes, indicating it is over temperature during the recording process.


Button	Indicator light	Status
Press and hold the “▶” button for 3s for the first time.	The “OK” (green light) is lighting for 3s.	Start Recording
After turning on, press and hold the “■” button for 3s.	The “Alarm” (red light) is lighting for 3s.	Stop Recording

## Battery Status Indication


Battery Status Indication	Capacity
	40% ~ 100%
	15% ~ 40%
	5% ~ 15%
	< 5%

**Note:** When the battery power is lower than 5%, please replace the CR2032 battery with a new one.


## Battery Replacement




Step 1




Step 2




Step 3




Step 4




Step 5



Step 6



Step 7



Step 8

**Step1:** Turn the battery cover counterclockwise to open it.

**Step2:** The battery cover position is as shown, indicating that it has been rotated to the accordant position.

**Step3:** Take off the battery cover.

**Step4:** Remove the CR2032 battery.

**Step5:** Replace the battery. (Please note that the positive and negative anodes)

**Step6:** Please install the battery cover as the direction shown.

**Step7:** Cover the battery and clockwise rotate it to close.

**Step8:** The battery cover is oriented as shown, indicating that the battery cover is installed tightly.

### Note:

It is recommended to check the battery status before restarting the logger to ensure that the remaining battery capacity could finish the recording task.

When the logger is connected to computer under the status of recording or pause status, it is prohibited to plug out the logger without battery power supply.